


Synchronization

This section describes how synchronization is achieved.

- TDT data is stored as a continuous data (starting from start of recording)
- behavioral data is stored per trial (starting from first trial)

Synchronization is enabled by sending state and trial information from monkeypsych to TDT ([Online communication](#)). Both state and trial information is sent through the same 8 pins, a little trick allows using the same channel for both purposes ([State signal](#)). For combining both, we first convert TDT data to trial format (function `TDT_trial_struct.m`) preferentially using epochs store (Tnum and SVal) information. There is a flag in `TDT_trial_struct` for using the continuous state information stream (`stream_state_info`), but it is permanently turned off, as the epoch information is already extracted from the continuous state information within the TDT circuit (see [RCO](#)) Finally, when fitting TDT data into a trial structure, certain timing conventions have proved useful: [Trial structure timing](#)

Synchronization during recording



Note that pins are only switched either to send trial information or after a state change (AND after the flip happened!) The idea is that both trial information and state information can be sent via the same channel. Details are described below.

Resulting 8 pin state signal




Values of 252 to 255 are preserved for indicating start of trial/state information and in-between trial information stoppers:

- 252: indicates start of trial information
- 254: in between trial information to define trialinfo packages

- 253: end of trial information
- 255: start of state information

Note that due to preserved values, a trial information package maximum value is limited. For example, to encode the date first package would be first 2 digits of the year, second package the last 2 digits of the year, third package the month and fourth package the day. That way none of the packages reaches a value of over 100. Trial information is sent during INI (state 1), during the other states the stat value reflects the current state.

TDT Internal RCO



Snippet and Stream synchronization in the combined trial structure



Importantly, trial start reference is STATE 2 (fixation acquisition). Trial end is STATE 2 of the next trial. Therefore, ITI between trial 1 and 2 as well as STATE 1 (initiation) of trial 2 are found at the end of trial 1.

This also means that data recorded before 1st trial's state 2 can not be stored in the trial structure and is saved in a separate variable (First_trial_INI) instead. Everything before first trialsINI is discarded.

From:
<http://dag.dokuwiki.dpz.lokal/> - DAG wiki



Permanent link:
http://dag.dokuwiki.dpz.lokal/doku.php?id=ephys_pipeline:2_synchronization&rev=1647525990

Last update: **2022/12/29 07:15**